

Northern Health and Social Care Trust Board Meeting in Public - Cover Sheet

Subject:	Approval of addendum for Upgrade of Endoscopy Decontamination Equipment and Ventilation systems in Antrim Hospital Day Procedure Unit		Date of Meeting: 27 th November 2025	
Prepared By:	Brian Boyd Trust Decontamination Manager			
Approved By:	Trust Board			
Presented By:	Dr George Gardiner			
Purpose				
To provide a summary of the addendum required for the upgrade of Endoscopy Decontamination Equipment and Ventilation systems in Antrim Hospital Day Procedure Unit for Trust Board approval.			Approval	✓
			Assurance	
			Update	
			Noting	
Strategic Objectives				
Build Northern Partnerships and Integrate Care	Continue to improve Outcomes & Experience	Deliver value by optimising Resources	Deliver value by optimising Resources	Improve population Health & address health & social care inequalities
	✓	✓	✓	
Risk – please note if links to any principal, corporate or divisional risk				
	Without investment into Antrim Day Procedure Unit there was an increased risk of – •Terminal failure of Endoscope Decontamination Equipment resulting in no Endoscopes being processed and procedures potentially being cancelled. •Failure to comply with HTM01-06, HTM 03-01 & BSEN16442:2015 •Potential adverse effect on patient safety •Non-compliance to HTM03-01			
Committees/groups where this item has been presented before				
Medical Device and Decontamination Committee				
Acronyms				
N/A				
Executive Summary				
Original Approval Value: £1,448,197 (24/11/22) + (Previous Addendums) £48,629 (30/05/24) + £188,000 (22/10/24) + £1,500 (25/11/24) + £3,000 (12/12/24) + £77,500 15/01/25 ----- Please note project now completed Current Addendum to allocation -BD230048 £22,717 for approval This current addendum is required as a result of a number of compensation events being negotiated since project completion and should get us to final account on the project The purpose of the original business case approved was to ensure the continuity of Endoscopy services on the Antrim Hospital site by providing clinical staff with the appropriate clean equipment for procedures. This				

We provide compassionate care with our community, in our community

Trust Board Reference Number: TB168/16/L

will ensure staff and clients of Antrim Hospital Day Procedure Unit are provided with a safe, comfortable and compliant environment.

Replacement of the decontamination equipment will ensure that the machines are compliant reliable and with manufacturers support to ensure the Endoscopy service can continue at full capacity.

Introduction of a 30-day vacuum pack storage system will negate the need for reprocessing of endoscopes after they have reached their 7 day expiry.

Replacement of the Endoscope Washer Disinfectors, Drying Cabinets, Upgrade of the Ventilation system and a redecoration throughout ensured compliance with HTM01-06, HTM03-01 & BSEN16442:2015